

Q. No.	Question	Response
1	Site have provided an overall list of assets along with maximum number of days, can we assume that multiple engineers can attend site for each month as some of the visits would require this in order to complete the full works within the maximum number of days?	Normally each visit would be 1 engineer, the exception being our Jenner building which has 2 engineers for the service visit.
2	Reading the spec we also would assume that KI testing would be require each visit, can you confirm this?	Yes a KI discuss is to be performed every visit
3	In order to allocate the correct number of engineers we would require the sizes of the safety cabinets	The majority of our safety cabinets are a single KI discuss test (1200mm) with 6 safety cabinets over 1500mm triple KI Discuss test
4	Are you aware of which areas are restricted as fumigation of the equipment is required out of these areas and a return to site the following day to collect. Alternatively due to the number of visits required it could be that equipment is left on site, we would need to take this into account in the pricing but may be worth a discussion.	For all areas where equipment is required to be fumigated out of the building TPI will provide 1 complete set of test equipment.
5	We do not carry a waste carrier license so any waste from site such as filters would need to be disposed of through site waste streams	TPI would dispose of HEPA filters and a small amount of general waste for this contract.
6	Where the email response is 24 hours please note the service desk is only manned Monday-Friday	If an email is sent on a Friday the next working day would be acceptable.
7	Where it notes the scope for PCR units it notes that this is to be compared to design requirements of 12469, we just wanted to check if this is a typo as above this is noted 14644-3 for the PCR units."	Yes this is a mistake 14644-3 is correct.
8	Please could you confirm the aperture width for all MBSCs covered within this tender?	The majority of our safety cabinets are a single KI discuss test (1200mm) with 6 safety cabinets over 1500mm triple KI Discuss test
9	Are you able to provide a list of the specific areas that are classified as 'Restricted'?	Our main building (Plowright) houses 102 safety cabinets all of which are in our restricted area. 1 additional safety cabinet is in our Biggs building which is also a restricted area. By the term "restricted" this is where we have a requirement to change into specialist clothing which is provided by TPI, on exiting these areas there is a requirement to shower out of the restricted area.
10	Are you able to provide a list of the specific areas that are classified as 'High Containment Laboratories'?	See above
11	Where MBSCs require 6 monthly attendances, would KI Operator Protection testing be carried out on both visits or just once per annum?	Yes a KI discuss is to be performed every visit
12	Would calibration certificates be made available for the equipment provided by TPI?	Yes calibration certificates can be made available.
13	Would it be possible to confirm which Building/Month the 6 >1500mm Safety Cabinets fall under?	1 off BSC - Plowright East Wing - Due January & June 1 off BSC - Plowright North Wing - Due March & September 1 off BSC - Plowright West Wing - Due April & October 3 off BSC - Jenner - Due February
14	So that we understand fully can you please provide a comprehensive asset list which specifically details the width of individual cabinets? We also need to know whether the "number of days available" ie eight in February are consecutive days Monday – Friday or hypothetically only Mondays and Tuesdays in February?	The majority of our safety cabinets are a single KI discuss test (1200mm) with 6 safety cabinets over 1500mm triple KI Discuss test. The number of days available are consecutive days.